

Veith Symposium 2024

Why Do IDEs Fail To Get FDA Approval: Updates To The FDA 510(k) Program And Why They Are Important

Valerie Merkle, PhD
Senior Director, Medical & Scientific Affairs - Terumo Aortic
vmerkle@terumoaortic.com




1

Disclosures

- Terumo Aortic employee
- Adjunct Professor at The University of Maryland
- US Device & Pharmaceutical Regulation
- Former FDA Lead Reviewer


Note: Views communicated are my own & not reflective of an official Terumo Aortic position



2

Agenda


- **Investigational Device Exemptions (IDEs)**
 - Overview of an IDE
 - Why do they fail to get approval?
- **510(k)s**
 - Program Overview
 - New FDA Guidances Available



3

Investigational Device Exemptions (IDEs)


- An IDE **allows the clinical investigation** of medical devices in the United States.
- Per 21 CFR 812.1: "An approved investigational device exemption (IDE) **permits a device** that otherwise would be required to comply with a performance standard or to have premarket approval **to be shipped lawfully for the purpose of conducting investigations of that device**."
- **In summary:** It's the means by which a manufacturer or physician is able to clinically investigate a new medical device (unapproved) or new use of an approved device in the United States
 - e.g., Early Feasibility Study, Feasibility Study, Pivotal Study
 - e.g., physician modification of a commercial device, new ascending endovascular graft or an expanded indication for a TEVAR device (for example, transection)



4

Reasons for Disapproval


- Reasons per 21 CFR 812.30(b) & Section 520 of FD&C Act
- **Failure to comply with the requirements** of 21 CFR 812, Section 520 of FD&C Act, or any other applicable regulation
- Application has an **untrue statement of material fact or omits information** required by 21 CFR 812
- Sponsor **fails to respond to request** for additional information in the time that FDA prescribes (21 CFR 812.30(b)(3))
- Reason to believe that **risks to subjects are not outweighed by anticipated benefits** to the subjects & the importance of the knowledge to be gained (21 CFR 812.30(b)(4))
 - Inadequate safety in the investigational plan
- **Inadequate informed consent** (21 CFR 812.30(b)(4))
- **Scientifically unsound** clinical investigation (21 CFR 812.30(b)(4))
- Reason to believe that the **device will be ineffective** (21 CFR 812.30(b)(4))
 - Inadequate potential benefit



5

Most Common Reasons for IDE Disapproval

- **Inadequate Report of Prior Investigations**
- **Missing or inadequate information** (e.g., animal study, engineering testing, biocompatibility evaluation)
- **Inadequate sample size and/or sample dimensions** evaluated to provide confidence/reliability on performance aspects of the device in the intended patient population
- **Inadequate clinical mitigations** for any risks that can't be addressed via testing (e.g., animal, engineering, other)
- **Insufficient discussion** as to how the **benefits outweigh the risks** for the intended patient population
- **Inadequate discussion as to how the totality of the data/information provided supports that the benefits outweigh the risks** for the intended patient population



6

Study Design


- FDASIA Section 601 - Signed into Law July 9, 2012
- **FDA shall not disapprove an IDE on the basis of FDA's belief that the study design is inadequate to support a future premarket approval (PMA) application, 510(k), Humanitarian Device Exemption (HDE) or De Novo Classification**
- **NOTE:** Concerns with the Study Design are communicated as Study Design Considerations (SDCs) and are **NOT** reasons for disapproval
 - SDC need to be taken seriously and addressed in a timely manner



7

510(k) Premarket Notification

- **510(k) Premarket Notification**
 - "premarket submission that is intended to demonstrate that the device to be marketed is **at least as safe and effective as legally marketed device** that does not require PMA."
 - Section 510(k) of the Food, Drug, and Cosmetic Act
 - Totality of the information must show **substantial equivalence**
- **Marketing Submission Pathway for many Class II devices**, e.g.,
 - Many balloon catheters (non-drug coated)
 - Vascular embolization devices (control hemorrhaging due to aneurysms, certain types of tumors (e.g., nephroma, hepatoma, uterine fibroids), and arteriovenous malformations)




8

510(k) Guidance

No change to applicable statutory & regulatory standards; New guidances are intended to improve transparency, predictability, & consistency.

- 510(k) submission documentation should show that **subject device is as safe & effective as the predicate device & demonstrate substantial equivalence (SE)**
- **Predicate Device:** A legally marketed device to which the subject device of a 510(k) submission is compared to
- **FDA Guidance:** Best Practices for Selecting a Predicate Device to Support a Premarket Notification 510(k) Submission
- Select predicate device that was cleared using well-established methods (e.g., FDA recognized voluntary consensus standards, FDA guidances)
- Predicate devices meet or exceed expected safety & performance
 - Consider device's safety and/or effectiveness information (e.g., adverse events/complaints) after being commercially distributed
- Predicate devices without unmitigated use-related or design-related safety issues
 - Consider whether there are any emerging signals (causal association or new aspect of known association w/AE) and/or safety communications
- Predicate devices without an associated design-related recall




9

510(k) Guidance

No change to applicable statutory & regulatory standards; New guidances are intended to improve transparency, predictability, & consistency.


- **In many cases:** substantial equivalence is demonstrated through robust non-clinical safety and performance data (no clinical data needed)
 - e.g., technological characteristics & intended use are same/sufficiently similar as predicate device
 - Clinical data may be needed to show substantial equivalence (SE) to the predicate device
- **FDA Guidance:** Recommendations for the Use of Clinical Data in Premarket Notification 510(k) Submissions
- **Four Potential Scenarios**
 - Differences in the Indications for use of the new device & predicate device
 - Differences in technological characteristics of the new device & predicate device
 - SE b/t new device & predicate device cannot be determined by non-clinical testing (e.g., analytical, bench, animal)
 - Newly identified or increased risk for predicate device suggests clinical data may be needed for new device



10

Summary


- IDEs are generally disapproved due to Inadequate Report of Prior Investigations, e.g., Incomplete testing, insufficient discussion as to the totality of the data supports that the benefit outweighs the risks
- New FDA Guidances aim to further clarify the 510(k) process
- **FDA Guidance:** Recommendations for the Use of Clinical Data in Premarket Notification 510(k) Submissions
- **FDA Guidance:** Best Practices for Selecting a Predicate Device to Support a Premarket Notification 510(k) Submission
- **Note** - No change to applicable statutory & regulatory standards; New guidances are intended to improve transparency, predictability, & consistency.



11

Thank you!

Valerie Merkle, PhD
Senior Director, Medical & Scientific Affairs - Terumo Aortic
vmerkle@terumoaortic.com



12